

---

# NI-9853 and sbRIO-9853 Specifications

---

2024-07-24



# Contents

NI-9853 and sbRIO-9853 Specifications..... 3

# NI-9853 and sbRIO-9853 Specifications

## Definitions

**Warranted** specifications describe the performance of a model under stated operating conditions and are covered by the model warranty.

**Characteristics** describe values that are relevant to the use of the model under stated operating conditions but are not covered by the model warranty.

- **Typical** specifications describe the performance met by a majority of models.
- **Nominal** specifications describe an attribute that is based on design, conformance testing, or supplemental testing.

Specifications are **Typical** unless otherwise noted.

### Related information:

- [Software Support for CompactRIO, CompactDAQ, Single-Board RIO, R Series, and EtherCAT](#)

## Conditions

Specifications are valid for the range -40 °C to 70 °C unless otherwise noted.

## High-Speed CAN Characteristics

Transceiver	Philips TJA 1041
Max baud rate	1 Mbps

CAN_H, CAN_L bus lines voltage	-27 V DC to +40 V DC
<b>Supply voltage range (V sup)</b>	
CAN1	+8 V DC to +25 V DC
CAN0	N/A
MTBF	1,816,913 hours at 25 °C; Bellcore Issue 6, Method 1, Case 3, Limited Part Stress Method



**Note** Contact NI for Bellcore MTBF specifications at other temperatures or for MIL-HDBK-217F specifications.

## Safety Voltages

Connect only voltages that are within the following limits:

<b>Maximum Voltage<sup>1</sup></b>	
Port-to-COM	-27 V DC to +40 V DC maximum, Measurement Category I
External power supply (Vsup)	+30 V DC
<b>Isolation</b>	
<b>Port-to-port</b>	

1. The maximum voltage that can be applied or output between any port or V<sub>SUP</sub> terminal and a COM terminal without creating a safety hazard.

Continuous	60 V DC, Measurement Category I up to 5,000 m altitude
Withstand	500 V RMS, verified by a 5 s dielectric withstand test
<b>Port-to-earth ground</b>	
Continuous	60 V DC, Measurement Category I up to 5,000 m altitude
Withstand	500 V RMS, verified by a 5 s dielectric withstand test

## Measurement Category I



**Warning** Do not connect the product to signals or use for measurements within Measurement Categories II, III, or IV, or for measurements on MAINS circuits or on circuits derived from Overvoltage Category II, III, or IV which may have transient overvoltages above what the product can withstand. The product must not be connected to circuits that have a maximum voltage above the continuous working voltage, relative to earth or to other channels, or this could damage and defeat the insulation. The product can only withstand transients up to the transient overvoltage rating without breakdown or damage to the insulation. An analysis of the working voltages, loop impedances, temporary overvoltages, and transient overvoltages in the system must be conducted prior to making measurements.



**Mise en garde** Ne pas connecter le produit à des signaux dans les catégories de mesure II, III ou IV et ne pas l'utiliser pour des mesures dans ces catégories, ou des mesures sur secteur ou sur des circuits dérivés de surtensions de catégorie II, III ou IV pouvant présenter des surtensions transitoires supérieures à ce que le produit peut supporter. Le produit ne doit pas être raccordé à des circuits ayant une tension maximale supérieure à la tension de fonctionnement continu, par rapport à la terre ou à d'autres voies, sous peine d'endommager et de compromettre l'isolation. Le produit peut

tomber en panne et son isolation risque d'être endommagée si les tensions transitoires dépassent la surtension transitoire nominale. Une analyse des tensions de fonctionnement, des impédances de boucle, des surtensions temporaires et des surtensions transitoires dans le système doit être effectuée avant de procéder à des mesures.

Measurement Category I is for measurements performed on circuits not directly connected to the electrical distribution system referred to as **MAINS** voltage. MAINS is a hazardous live electrical supply system that powers equipment. This category is for measurements of voltages from specially protected secondary circuits. Such voltage measurements include signal levels, special equipment, limited-energy parts of equipment, circuits powered by regulated low-voltage sources, and electronics.



**Note** Measurement Categories CAT I and CAT O are equivalent. These test and measurement circuits are for other circuits not intended for direct connection to the MAINS building installations of Measurement Categories CAT II, CAT III, or CAT IV.

## Environmental Characteristics

<b>Temperature</b>	
Operating	-40 °C to 70 °C
Storage	-40 °C to 85 °C
<b>Humidity</b>	
Operating	10% RH to 90% RH, noncondensing
Storage	5% RH to 95% RH, noncondensing

Ingress protection	IP40
Pollution Degree	2
Maximum altitude	5,000 m
<b>Shock and Vibration</b>	
<b>Operating vibration</b>	
Random	5 g RMS, 10 Hz to 500 Hz
Sinusoidal	5 g, 10 Hz to 500 Hz
Operating shock	30 g, 11 ms half sine; 50 g, 3 ms half sine; 18 shocks at 6 orientations

To meet these shock and vibration specifications, you must panel mount the system.

## Power Requirements

<b>Power consumption from chassis</b>	
<b>Active Mode</b>	
Transmitting	625 mW maximum
Receiving	400 mW maximum
Sleep mode	25 $\mu$ W maximum
<b>Thermal dissipation (at 70 °C)</b>	

Active mode	1 W maximum
Sleep mode	250 mW maximum

## Physical Characteristics

Dimensions	Visit <a href="https://ni.com/dimensions">ni.com/dimensions</a> and search by module number.
Weight	Approximately 144 g (5.0 oz)